

Specification مواصفات

CODE NO	ITEM NO. رقم البند	ITEM DESCRIPTION اسم البند
	A003	ANAESTHESIA, SUPPLY CART عربة تخدير

Manufacturer الشركة الصانعة	Country of Origin بلد المنشأ	Model الموديل

Parameter Description عنصر المواصفة	Specified المطلوب	Offered المعروض
--	----------------------	--------------------

1	ANAESTHESIA, SUPPLY CART		
	1	Mobile unit, impact resistant plastic top with raised edges.	Drawers unit with large and small drawers
	2	Swing out side storage unit with tray divider.	Sharps container holder, and tilt out bin.
	3	Multi-purpose overbridge unit .	With universal clamp, blade /bag holder, small. large metal bin, label/tape dispenser, & bins w/ cover.
	2	COMPLETE WITH THE FOLLOWING:	
	1	Forceps, Magil, Adult.	set qty 2 set
	2	Forceps, Magil, Child.	set qty 2 set
	3	All needed accessories.	Specify
	3	OTHER SPECIFICATION	FDA, CE, ISO Approved

<p>Important Note: Contractor must fill the "Offered" column and submit this table as a part of the tender document. The number of the related parameters specified in the specification shall be clearly marked /shown on the catalogue. Non compliance with this note will void the offer.</p>	<p>ملاحظة هامة: يجب على مقدم العطاء تعبئة خانة المعروض وتقديم هذا الجدول ضمن اوراق المنافسة بعد ختمه وتوقيعه، وتكتب ارقام عناصر المواصفة على ما يقابلها على الكatalog، ولن ينظر لعرضه ما لم يلتزم بذلك</p>
<p>The fixation of the equipment data plate from factory must be permanent and non-removable, and the data shown on it must be embossed/engraved or printed permanently, that guarantees the data cannot be altered.</p>	<p>ملاحظة هامة: يجب أن تكون اللوحة الاسمية للجهاز مثبتة على الجهاز بشكل دائم وغير قابلة للإزالة، كما يجب أن تكون المعلومات المدرجة فيها محفورة أو مطبوعة بشكل غير قابل للإزالة أو التغيير.</p>